

KERALA VETERINARY & ANIMAL SCIENCES UNIVERSITY
TEACHING VETERINARY CLINICAL COMPLEX

Pookode, Lakkidi (PO), Wayanad 673 576
Phone: 8547144462 E-mail: tvccpkt@kvasu.ac.in

No. TVCC/ Pkt/ WDP 23-24/2025

Date: 06-01-2026

TECHNICAL BID AND DATE EXTENSION CORRIGENDUM

Tender ID : 2025_KVASU_823328_1
Tender reference Number : TVCC/ PKT/ WDP 23-24/2025/
eTender1 (R)

Competent authority has made some changes (Amendments) in respect of the aforementioned tender in terms of date extension and technical specifications. The amendments are detailed as below. It is requested to quote the offers along with compliance of the changes made to this tender.

1. Date Extension:

Last date of Submission of tenders: 13-01-2026 05.00 pm
Date of Opening of technical BID : 15-01-2026 10.00 am

2. Technical Bid modification:

Sl. No	Sl. No in the Tender Notice	Current Specification	Modified as
1	5.d	Bucky should have facility for Automatic Exposure Control.	Both horizontal and vertical buckies should have a facility for Automatic Exposure Control.
2	8.20	Detector should have facility for storing up to 200 images.	The detector should have a facility for storing up to 100 images or more
3	8.9	MTF should be 98% or more at 0 cyc/mm at RQA 5 beam or $80 \pm 10\%$ at 0 cyc/mm	Cancelled
4	10.1	The system must be a Dry Laser Imager	The system must be a Dry Laser Imager/ Thermal printer

Yours faithfully,

Professor and Head i/c



PROFESSOR & HEAD
Teaching Veterinary Clinical Complex
Pookode, Lakkidi (PO), Wayanad
PIN: 673 576